



FTT; THP R/L - steam

Machine-Type:	Tray Washer		Heating:	Steam
Model:	FTT	FTT; THP	Operation:	Right / Left
Usable-Width:	500	Usable-Height:	40	Main-Switch: Built in Machine
Chemical	conduit for chemical supply			Dimension: DN50, Position in mm: 100mm AFFL
Heating	Consumption		Dimension: DN25	Connection: G1" male, Position in mm: 50mm AFFL
5.1	Condensate	23,5 kg/h	49037 kJ/h	50mm AFFL
5.0	Steam			100mm AFFL
Steam-Flow-Pressure provided by customer: 0,5-1,0 bar				
Exhaust	Volume	Temp.	Rel. Humidity	Pressure
4.0	350 m³/h	18-22°C	90-98%	ca. 0 Pa
Exhaust	Control and Data-Line			Extended-Length
3.7	Equipotential	min. 1x6mm² provided by customer		3m reserve
3.5	Malfunction-Sensor		5x1,5 mm²	3m reserve
3.4	Exhaust		3x1,5 mm²	3m reserve
3.3	Dosage-System		7x1,5 mm²	3m reserve
Exhaust	Voltage	Frequency	Supply	Fuse
3.0	400 V	50 HZ	3-N-PE	3x35 A
				Cross-Section: 5x6 mm², Power: 13,5 kW, Extended-Length: 3m reserve
Water	Consumption	Temp.	Hardness	Conductance
1.1	105 l (Filling)	50-60 °C	max. 7°d (1,2mmol/l)	150-400µS/cm
1.0	200 l/h	12 °C	0-3°d (0,5mmol/l)	80-120µS/cm
Water-Flow-Pressure provided by customer: min. 1,5 bar / 22 psi				
Heat-Radiation (thermal output to the room)				
	washware	5,8 kW	latent	1,3 kW
			sensibel	4,6 kW

HOBART GENERAL INFORMATION

Installation: All installations should always comply with all national and local codes of practice.

Exhaust: A frost-protection flap is recommended if the exhaust air from the machine is exterior ducted.

Transport: Minimum measurements of entry doors for machine assembly = outer measurements of largest machine + 300mm in height, + 400mm in width!

Aeration: The ventilation and exhaust for the room must be interpreted according to local by-laws

Heat pump: Proper working order can only be ensured by room temperature above 18°C.

Shut-off valves: The shut-off valves for rinse water, tank filling or demi-rinse are supplied by others.

Control- and data lines: We recommend a conduit for control-lines in the area of the electrical connection.

Washing result: A spotless cleaning results can be achieved only with low mineral content of the rinse water. We recommend a conductance of about 80µS/cm.

Floor drain: Splash floor drains should be installed for general cleaning purpose.

Steam-connections: The in-house steam-flow piping must be equipped with condensate drainage prior to the HOBART steam-connection. The in-house condensate piping must be pressure less and able to absorb all HOBART condensate.

HOBART GENERAL LEGEND

AW = drain water (CNS)	KW = cold water	üOKFF = above finished floor
Dat = dataline	KWw = cold water soft	UK = lower edge
EZ = power line 230V / 400V	LR = conduit Ø	VEW = demineralized water
FD = floor opening	CNS = stainless steel (inox)	WD = wall opening
HW-VL = hot water flow	MK = supply channel	WS = wall slot
HW-RL = hot water return	PE = equipotential conductor	WW = warm water
ND = low pressure steam	STL = control line	WWw = warm water soft
Kond. = condensate	KB = cored hole Ø	

HOBART		HOBART GmbH - Robert-Bosch-Straße17 - 77656 Offenburg, Germany		Angebot <input checked="" type="checkbox"/> Quotation	
		Tel.: +49(0)781.600-0 - Fax.: +49(0)781.600-2319 - www.hobart.de		Auftrag <input type="checkbox"/> Order	
Datum / Date:	2012 / AUGUST	Project:	TRAY WASHER FTT HEATPUMP THP STEAM HEATED		
Gezeichnet / Drawn by:	WH	Geprüft / Checked by:	W: Neumaier		
Projectmanager:	Maßstab / Scale:	Order-No.:	Zeichnungsnummer / Drawing-No.:		
	1:20		36 FTT THP - steam RL		